



**广 微 测**  
Gmicro Testing



201819000883



中国认可  
国际互认  
检测  
TESTING  
CNAS L1747

# GUANGDONG DETECTION CENTER OF MICROBIOLOGY

## REPORT FOR ANALYSIS

**Report №.**

2020SP9964R01E

**Name of Sample**

Clean light sterilization wand

BINGRI ELECTRONIC TECHNOLOGY (SHENZHEN)

**Applicant**

CO.,LTD.

**Test Type**

Entrustment Test



**Address:** Building 66, No.100 Central Xian Lie Road, Guangzhou, China

**Postcode:** 510070

**Tel:** +86 20 87137666

**Fax :** +86 20 87137668

**Website :** www.gddcm.com



**广微测**  
Gmicro Testing



201819000883



中国认可  
国际互认  
检测  
TESTING  
CNAS L1747

## GUANGDONG DETECTION CENTER OF MICROBIOLOGY

### REPORT FOR ANALYSIS

Report №.: 2020SP9964R01E Verification Code: 34075926

Name of Sample	Clean light sterilization wand	Test Type	Entrustment Test
Applicant	BINGRI ELECTRONIC TECHNOLOGY (SHENZHEN) CO.,LTD.	Address	B3 Merchants Guangming Science Park, Guangguang Road, Guangming District, Shenzhen
Sample Source	Submitted for Testing by the Applicant	Sample Quantity	One
Spec and Lot № of Sample	—	State and Characteristic	Machine
Sample Received Date	2020-09-10	Test Completion Date	2020-09-17
Test Standard and Method	Technical Standard for Disinfection (Ministry of Health, Edition 2002), section 2-2.1.5.4		
Item Tested	Sterilization and disinfection equipment disinfection efficacy test: UV lamp		
Test Conclusion	The test data of the sample(s) is attached to the page(s) of this report.		
Remarks	Manufacturer: BINGRI ELECTRONIC TECHNOLOGY (SHENZHEN) CO.,LTD. (Provided by the applicant)		

Issue Date: 2020-09-18

(Official Seal)



Editor: *Liyunqian*

Verifier: *Liang Daqing*

Approver: *Gao Weipeng*



GUANGDONG DETECTION CENTER OF MICROBIOLOGY


ANALYSIS AND TEST RESULT

Report No.: 2020SP9964R01E

- 1. Action Time: 1s
- 2. Action distance: 1cm
- 3. Results:

Test organisms	Test group	Average cfu of testing groups (cfu/ piece)	Average cfu of positive controls (cfu/ piece)	Sterilization logarithm (KL)	Sterilization rate (%)
<i>Escherichia coli</i> 8099	1	<5	$9.0 \times 10^5$	>5.25	>99.99
	2	<5	$8.0 \times 10^5$	>5.20	>99.99
	3	<5	$9.0 \times 10^5$	>5.25	>99.99
	Average			>5.23	>99.99
<i>Staphylococcus aureus</i> ATCC 6538	1	<5	$2.1 \times 10^6$	>5.62	>99.99
	2	<5	$2.4 \times 10^6$	>5.68	>99.99
	3	<5	$2.7 \times 10^6$	>5.73	>99.99
	Average			>5.67	>99.99
<i>Staphylococcus albus</i> 8032	1	<5	$1.7 \times 10^6$	>5.53	>99.99
	2	<5	$1.9 \times 10^6$	>5.58	>99.99
	3	<5	$1.7 \times 10^6$	>5.53	>99.99
	Average			>5.55	>99.99
<i>Candida albicans</i> ATCC 10231	1	50	$2.5 \times 10^6$	4.70	99.99
	2	40	$2.4 \times 10^6$	4.78	99.99
	3	25	$2.2 \times 10^6$	4.94	99.99
	Average			4.81	99.99

Logo:



(Blank Below)

分析  
专用



Report №.: 2020SP9964R01E

## Notice Items

1. The Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.
2. The Test report is invalid without signature of verifier and approver.
3. The Test report is invalid if being supplemented, deleted or altered.
4. Without prior written permission, the report cannot be reproduced, except in full.
5. Unless otherwise stated, the results shown in this test report refer only to the sample(s) submitted.
6. Any dispute of the report must be raised to the testing body within 15 days after the report is received, exceeding which the dispute will not be accepted.
7. For the tested sample(s) submitted by the applicant, the sample information in the test report is provided by the applicant and the laboratory is not responsible for its authenticity.

